


Reducing Sugars

Example 1 - Range: 20 - 400 mg/L



Reducing Sugars A041-A00

Range: 20 - -400 mg/L

Mean=N/A

SD=N/A

RSD=0.87% at 20 mg/L

MDL=N/A

Analysis rate: 72 samples/hour